

- 42 -

## ABSTRACT OF THE DISCLOSURE

In a nuclear magnetic resonance spectrometer, the shape of a detection coil is changed from a conventional cage type to a solenoid type of higher sensitivity. Accordingly, differing from the conventional superconductive magnet of multilayer air core solenoids, a superconductive magnet is right and left divided to split magnets for generating 11 T, preferably, 14.1 T in the horizontal direction, and the magnetic field uniformity is set to 0.001 ppm or less and the temporal stability is set to 0.001 ppm or less.